




Bottle Top Dispenser Valve



Mole_NLD

[VIEW IN BROWSER](#)

updated 9. 10. 2026 | published 9. 10. 2026

Summary

Replacement valve for lab bottle top dispensers like the "Dispensette"



0.21 hrs



1 pcs



0.15 mm



0.40 mm



PET



1 g



Prusa CORE
One L

[Healthcare](#) > [Medical Tools](#)

Tags: [valve](#) [tap](#) [dispenser](#) [brand](#) [valvehandle](#)
[dispensette](#)

One of our valves broke off from our Brand Dispensette bottle top dispensers. As I could not find a replacement readily available I decided to sit down, draw a bit and this came out. Perfectly snug fit, and with a very subtle fuzzy skin it's grippy too! Note the modifiers in the printfile has 2 modifiers so only the fingerholds are grippy. You want the portion in the dispenser to be smooth for ease of fit and operation.

I added my rough sketches and measurements in the photos as well for those interested.

Model files



dispensette-valve.f3d



dispensette-valve.step



dispensette-valve.3mf

Print files



dispensette-valve_04n_015mm_petg_coreonel_12m.bgcode

⊗ PET ⊕ 0.40 mm ≡ 0.15 mm ⌚ 0.21 hrs ⚖ 1 g 🖨 Prusa CORE One L

License ©

This work is licensed under the

Open Community License v1.1 + Research & Development v1 (OCL v1.1 + RnD v1)

- ✓ | Remix culture allowed
- ✓ | Patent license included
- ✓ | Right-to-repair protected
- ✓ | Internal commercial use
- ✗ | Commercial reselling
- i | Share under the same license
- i | No attribution necessary
- i | Research & Development only

